



Product Service

# Attestation of Conformity

No. N8 12 11 64589 004

**Holder of Certificate:** Esco Micro Pte. Ltd.21 Changi South Street 1  
Singapore 486777  
SINGAPORE**Product:** Laboratory Equipment  
Aeris Thermal cycler**Model(s):** AERIS-BG096, AERIS-B4830, AERIS-BG384,  
AERIS-BD048, AERIS-B4076**Parameters:**

Rated voltage:	100-240V~
Rated frequency:	50/60Hz
Rated input power:	600W
Protection class:	Class I
Degree of protection:	IP20

**Tested according to:** EN 61010-2-010:2003

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

**Test report no.:** 704201285501-00**Date,** 2012-12-05

( Junjie Sun )



After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 1