

invent-uk

ANSI/ASHRAE 110-1995

Flow Containment & Instrumentation

CERTIFICATE OF TYPE TESTING IN ACCORDANCE WITH ASHRAE 110-1995

CERTIFICATE & REPORT NO: INV/ASHRAE110/607

DATE: 30th November 2012



Fume Hood Manufacturer:

Esco Micro Pte Ltd
21 Changi South Street 1
Singapore 486 777

Fume Hood Type:

1.5m wide EFD-5B1, EFD-5B2, EFD-5B3 bench-type

Sash Opening Width: 1300mm

Sash Opening Height: 440mm (100%)

Average Face Velocity: 0.52m/s (at 100%)

This is to certify that the fume hood described above has been type-tested in accordance with ASHRAE 110-1995 and resulted in performance characteristics given in the corresponding test report.

Tested and Certified by: Dr A F Bicen

Invent UK Ltd, 85 Southdown Road, Harpenden, Herts AL5 1PR

Tel: 01582 461666 : Fax: 01582 460077 : E-mail: info@invent-uk.com

A handwritten signature in blue ink that reads 'A.F. Bicen'.