

Test-report**No.: 160312**

Customer: Esco Micro Pte. Ltd.
21 Changi South Street 1
Singapore 486777

Location: biomedis Kalibrierservice GmbH & Co. KG
Kerkrader Straße 2
D-35394 Gießen, Germany
Prüflabor 3

Object: Isotherm Laboratory Incubator

Manufacturer: Esco

Type: IFA 110-8

Serial no.: 2011-56950


biomedis ID-no.: --

Measuring procedure: The calibration was performed in accordance to ISO DIN 12880. 27 thermocouples were arranged in the chamber as you can see in the appendix "arrangement of sensors in the chamber". The useable volume was performed under unloaded condition.

Actual state: The device is calibrated and ready for use

Hereby we confirm that the calibration was realized by an ISO/IEC 17025 accredited testing laboratory. The measuring installations used for calibration are regularly calibrated and traceable to the national standards of the German Federal Physical Technical Institute (PTB).

Seal



Date of calibration

2012-03-16

Signature

(Thomas Angerer)

Date of the recommended re-calibration: 2013-03

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal are not valid. The test results refers only to the checked device.