



Biosafety Certificate No. 54-07 A



Esco Airstream Class II (AC2-6E1) Biohazard Safety Cabinet

Tests to BS EN 12469:2000, Biotechnology – Performance criteria
for microbiological safety cabinets

Customer Ref.	C-10301
Tests carried out for	Mr Alex Atmadi of ESCO
Report Number	54-07 A
Report Date	4 th October 2007

The Biosafety Unit of the Health Protection Agency Centre for Emergency Preparedness and Response, has undertaken extensive tests on the Esco Airstream Class II (AC2-6E1) biohazard safety cabinet. The cabinet was found to meet the performance required by BS EN 12469:2000 for all of the tests below.

- (A) Air velocity measurements.
- (B) Operator protection/containment tests using KI.
- (C) Operator protection tests using bacterial spores
- (D) Product protection tests using bacterial spores
- (E) Cross contamination tests using bacterial spores
- (F) Air visualisation tests.
- (G) Filter integrity tests.
- (H) Maintenance of temperature by the working cabinet.
- (I) Acceptable noise and light levels during operation of the cabinet.
- (J) Acceptable design of filter mountings and plenum.

Testing Performed By  Mr. Simon Parks Senior Research Scientist	Authorised By  Mrs Sara Speight Senior Research Scientist
---	---