

CERTIFICATE OF COMPLIANCE

Certificate Number 20140802-E248249
Report Reference E248249-A20-UL
Issue Date 2014-AUGUST-02

Issued to: ESCO MICRO PTE LTD
21 CHANGI SOUTH ST 1, SINGAPORE 486777
SINGAPORE


This is to certify that representative samples of LABORATORY USE ELECTRICAL EQUIPMENT
See Addendum Page

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1, (ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - Part 1: General Requirements), CAN/CSA-C22.2 No. 61010-1, (ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - Part 1: General Requirements)

Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Listing Mark for the US and Canada should be considered as being covered by UL's Listing and Follow-Up Service meeting the appropriate requirements for US and Canada.

The UL Listing Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers:  the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

Look for the UL Listing Mark on the product.



William R. Carney, Director, North American Certification Programs
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus



CERTIFICATE OF COMPLIANCE

Certificate Number 20140802-E248249
Report Reference E248249-A20-UL
Issue Date 2014-AUGUST-02

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Models:

Biological Safety Cabinet:

AC2 Model Series:

AC2-3S8, AC2-3S9, AC2-4S8, AC2-4S9, AC2-5S8, AC2-5S9, AC2-6S8, AC2-6S9, AC2-4N7, AC2-6N7, AC2-4Y7

AR2 Model Series:

AR2-4S8, AR2-6S8, AR2-4S9, AR2-6S9

FSC Model Series:

FSC-4S8 and FSC-4S9



William R. Carney, Director, North American Certification Programs
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

