



**BIU Certificate No. 875-03**



**Esco Labculture Flow Cabinet  
LHC-4A1**

**Tests to BS EN 12469:2000, Biotechnology –  
Performance criteria for microbiological safety  
cabinets**

Customer Ref.	EMS/32977
Issue Date	25 <sup>th</sup> August 2004
Copy Number	1
Tests carried out for	Mr X Q Lin of ESCO

The Esco Lab Culture Flow Cabinet (LHC-4A1) has passed the following tests using the methodologies specified in BS EN 12469:2000.

- A. Product protection tests using bacterial spores
- B. Cross contamination tests using bacterial spores
- C. Filter integrity tests

Testing Performed By  Mr. Simon Parks Research Scientist	Authorised By  Mr Allan Bennett Head of Biosafety
--	--