



Declaration of Certification

Esco Frontier Acela® Laboratory Fume Hood

Has been tested per the ASHRAE 110/2005 "Method of Testing Performance of Laboratory Fume Hoods"

And has received a rating of:

"AM" <0.01 ppm

At the design operational sash height of full open"

Testing Performed By:

Exposure Control Technologies, Inc.
231 C East Johnson Street Cary, NC 27513 USA
www.labhoodpro.com

January, 2010